

Certificate of Analysis

Hydrogen and Carbon Stable Isotope Ratio Reference Materials

n-Alkane Mixture Type A8

<i>n</i> -alkane	mean	one	range $\delta^2\text{H}$	n	mean	one	range $\delta^{13}\text{C}$	n	Concentrations of individual <i>n</i> -alkanes [mg in 0.5 mL]
	$\delta^2\text{H}$ ‰	σ ‰	‰		$\delta^{13}\text{C}$ ‰	σ ‰	‰		
C-16 #2	-9.1	1.4	-7.9 to -11.1	7	-26.15	0.02	-26.13 to -26.17	5	0.70
C-17 #2	-121.2	0.5	-120.6 to -122.0	5	-31.87	0.02	-31.84 to -31.90	8	0.70
C-18 #2	-52.0	1.1	-50.6 to -53.5	5	-32.70	0.01	-32.69 to -32.72	5	0.70
C-19 #2	-56.3	1.0	-55.0 to -57.5	5	-31.99	0.01	-31.98 to -32.02	6	0.70
C-20 #4	-49.6	2.1	-47.2 to -52.3	4	-31.87	0.02	-31.85 to -31.90	5	0.70
C-21 #2	-181.6	0.6	-180.7 to -182.3	5	-28.83	0.02	-28.81 to -28.85	5	0.70
C-22 #3	-68.2	1.8	-65.7 to -70.4	5	-34.89	0.02	-34.87 to -34.92	6	0.70
C-23 #2	-67.2	1.0	-65.6 to -68.6	6	-33.37	0.03	-33.33 to -33.40	5	0.70
C-24 #2	-29.7	1.5	-28.2 to -31.8	6	-32.13	0.02	-32.11 to -32.16	6	0.70
C-25 #4	-258.9	0.8	-258.1 to -260.0	5	-28.46	0.02	-28.42 to -28.48	7	0.70
C-26 #2	-45.9	1.0	-44.4 to -46.7	5	-32.94	0.01	-32.92 to -32.95	7	0.70
C-27 #4	-205.2	1.6	-203.5 to -207.6	6	-31.11	0.01	-31.11 to -31.12	5	0.70
C-28 #2	-36.8	1.3	-35.6 to -38.9	5	-33.20	0.01	-33.20 to -33.20	5	0.70
C-29 #5	-85.4	1.4	-82.9 to -86.8	6	-31.83	0.02	-31.80 to -31.85	5	0.70
C-30 #4	-41.5	0.7	-40.9 to -42.9	6	-33.14	0.02	-33.12 to -33.16	6	0.70


Notes

Commercially available individual pure *n*-alkanes were tested for purity of each compound using gas chromatography. Stable isotope ratios of individual compounds were determined at Indiana University by combusting multiple aliquots of each pure *n*-alkane off-line according to Schimmelmann et al. (1999, *Geochimica et Cosmochimica Acta* **63**, 3751-3766). Elemental hydrogen gases were measured and reported isotopically according to Coplen's (1996) guidelines relative to VSMOW2 (zero ‰) and normalized to SLAP2 (-427.5 ‰) (Coplen, T. B., 1996. New guidelines for reporting stable hydrogen, carbon, and oxygen isotope-ratio data. *Geochimica et Cosmochimica Acta* **60**, 3359-3360). Carbon isotope ratios are reported relative to the VPDB scale where NBS 19 and LSVEC are defined as exactly +1.95 and -46.6 ‰, respectively (Coplen et al., 2006. New guidelines for $\delta^{13}\text{C}$ measurements. *Analytical Chemistry* **78** (7), 2439-2441). Mass-spectrometric data in customary δ notation have a precision of ± 1.5 ‰ for $\delta^2\text{H}$ and ± 0.05 ‰ for $\delta^{13}\text{C}$. The vial of LSVEC used for calibration was rarely opened and thus not compromised.

The data in the Table shown above represent the mean value, one standard deviation, the range, and the number of analyses. Concentrations of each *n*-alkane in mixture A7 are indicated in terms of milligram *n*-alkane dissolved in 0.5 milliliter chromatographically pure hexane (Honeywell Burdick & Jackson, capillary GC GC-MS solvent). Aliquots of 0.5 mL of solution are sealed under argon in glass ampoules. Primary reference materials used for isotopic characterization were VSMOW2, SLAP2, NBS 19, and LSVEC. Isotopic and statistical analyses were performed at Indiana University by Arndt Schimmelmann.

Remedies for any claimed defect in these products will be limited to product replacement or refund of the purchase price. In no event shall Indiana University and their staff be liable for any damages, including but not limited to incidental or consequential damages. By accepting shipment of reference materials you agree to this limitation of liability.

This is a certified reference material (CRM) that can be traced to primary standards VSMOW, SLAP, NBS 19, and LSVEC.



CERTIFIED 3 September 2024

Arndt Schimmelmann, Ph.D., Senior Scientist, Indiana University, Department of Earth and Atmospheric Sciences,
1001 East 10th Street, Bloomington, Indiana 47405-1405, U.S.A, phone (812) 855-7645, e-mail: arndtschimme@gmail.com,
personal home page: <https://earth.indiana.edu/directory/faculty/schimmelmann-arndt.html>